

PLEASE NOTE

All users are requested to use following settings for luciferase assays :---

1. Amount of sample : **40 μ l**
2. Luminol/ sample : **25 μ l**
3. Time : **20 Sec**

Please don't forget to write down the **count** and **serial numbers** of vials used as well as the "total number of samples" tested in the Luminol Usasage record book placed at my (Rajeev) working place.